



Princeton BioMeditech Corporation

Makers of Advanced Rapid Diagnostic Products

P.O. Box 7139, Princeton, NJ 08543-7139, USA

Tel: 732-274-1000

4242 U.S. Highway 1, Monmouth Junction, NJ 08852-1905, USA

Fax: 732-274-1010

www.pbm.com / info@pbmc.com

November 20, 2025

MEMO

To: Customers of *Status*TM COVID-19/Flu A&B test kit

Re: Notification of Updated Expiration Dating

Dear Valued Customer,

The ***Status*TM COVID-19/Flu A&B test kit** ("Product") has recently received 510(k) clearance with CLIA-waived categorization, and the established shelf life of the Product has been extended from 24 months to 28 months based on real-time stability testing. The Product in your possession may now have an expiration date that is later than the labeled date. The affected Product lot numbers, originally labeled expiration date, and updated expiration date are listed below. All Products now have a maximum 28-month expiration date (subject to other kit components*). Please retain this memo with your Product records.

Lot Number	Printed Expiration Date	Updated Expiration Date
883H11	8/31/2025	12/31/2025
883J11	9/30/2025	1/31/2026
883J11A	9/30/2025	1/31/2026
883J11B	9/30/2025	1/31/2026
883J11C	9/30/2025	1/31/2026
883J21	9/30/2025	1/31/2026
883J21A	9/30/2025	1/31/2026
883K11	10/31/2025	2/28/2026
883K11A	10/31/2025	2/28/2026
883K11B	10/31/2025	2/28/2026
883L11	11/30/2025	3/31/2026
883L11A	11/30/2025	2/28/2026*
883L11B	11/30/2025	3/31/2026
883M11	12/31/2025	4/30/2026
884A11	1/31/2026	5/31/2026
884F11	6/30/2026	10/31/2026
884F11A	6/30/2026	10/31/2026
884F11B	6/30/2026	10/31/2026

884F21	6/30/2026	10/31/2026
884F21A	6/30/2026	10/31/2026
884F31	6/30/2026	10/31/2026
884F31A	6/30/2026	10/31/2026
884G11	7/31/2026	11/30/2026
884G11A	7/31/2026	11/30/2026
884G11B	7/31/2026	11/30/2026
884H21	8/31/2026	12/31/2026
884H21A	8/31/2026	12/31/2026
884H31	8/31/2026	12/31/2026
884H31A	8/31/2026	12/31/2026
884H31B	8/31/2026	12/31/2026
884J11	9/30/2026	1/31/2027
884J21A	9/30/2026	1/31/2027
884J21B	9/30/2026	1/31/2027
885A11	1/31/2027	5/31/2027
885H11	8/31/2027	12/31/2027
885H21	8/31/2027	12/31/2027
885H21A	8/31/2027	12/31/2027
885J11	9/30/2027	1/31/2028
885J21	9/30/2027	9/30/2027 *

We are pleased to communicate this improvement in the utility of the product and look forward to assisting you further. Please contact us if you have any questions regarding this memo. E-mail: techservice@pbmc.com

Sincerely,

Young-A Kim

 Young-A Kim, Ph.D.
 Director, QA/RA
 PBM



 Roger Kang
 President
 LifeSign, LLC